

EC Certificate

mdc medical device certification GmbH

Notified Body 0483
herewith grants

altona Diagnostics GmbH
Mörkenstraße 12
22767 Hamburg
Germany

for the scope

AltoStar® HCV RT-PCR Kit 1.5
(REF AS0211513)

the

EC Design Examination Certificate

The examination of the design of the product by mdc has proven
that the design meets the requirements according to

Annex IV – Section 4 of the Council Directive 98/79/EC

of the European Parliament and of the Council of
27 October 1998 on in vitro diagnostic medical devices.

This certificate is only valid in connection with a valid mdc certificate
according to Annex IV – excluding section 4 and 6 for the above mentioned products.

Valid from	2022-05-23
Valid until	2025-05-26
Registration no.	D1267700034
Report no.	P22-00749-236480
Stuttgart	2022-05-23


Head of Certification Body



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-247.10.05